IMPATH DS Block



Product Availability

Cat. No. 46174 Size 100 Tests PathCom Systems, Inc. 6759 Sierra Ct. Ste# B Dublin, CA 94568

Intended Use

For In Vitro Diagnostic Use.

Rev: 092813A

Description

ImPath DS Block is hydrogen peroxidase block that is intended for the antibodies that require the blocking of endogenious peroxidase prior to target retrieval of formalin-fixed, paraffin-embedded tissue sections in an immunohistochemistry (IHC) procedure in conjunction with ImPath 36.

Distributed by:

A.Menarini Diagnostics S.r.l. Via Sette Santi, 3 50131 Firenze Italy

Summary and Explanation

IHC (Immunohistochemistry) is a commonly used technique supporting the identification of antigens present on the tissue or cells.

It is the responsibility of the operator to identify the best working conditions and the best reagents to perform the staining run.

Via Gorizia 50 23900 Lecco Italy

Reagents Provided

Camananat	0
Component	Quantity per kit
DS Block	ı(one)





IMPATH DS Block



Dilution and Mixing

The product is ready-to-use. Number of tests is programmed with RFID on each vial. No further dilution is required.

Storage and Stability

Store at 2-8°C. Do not freeze. The reagents must be returned to the storage conditions identified above immediately after use. When properly stored, the reagents are stable to the date indicated on the label.

Warning and Precautions

ImPath Solutions are harmful and irritating to the eyes, respiratory system and skin. It may cause lung and stomach damage if ingested. Wear disposable gloves when handling reagents. MSDS are available from your local representative upon request. There are no express or implied warranties which extend beyond this datasheet. ImPath is not liable for personal injury, property damage or economic loss caused by this product.

Instruction for Use

Reagents for IHC procedures

Dilute (20x)wash buffer by mixing 19 portions of distilled water with 1 portion of the wash buffer to final 1x. Fill the wash buffer bottle from the instrument with the 1x wash buffer, screw the cap connected with tubing. Place the bottle in the designated area of the instrument.

Pull the Reagent Rack that has the required RFID tagged solution vials from refrigerator, remove the caps from the vials and place the Reagent Rack onto the ImPath 36. Turn on the unit, log in, press Prepare Labels to prepare the slide labels, place them on the slides, press Load Slides, put slides on the modules, press Scan Slides, then press Scan Reagents and allow the RFID Reader to identify and register the products, check on the required wash buffer and waste container, press Start Staining Process, or press Delayed Start and set the finished time you want, or click Return to Main Screen and press Start. The staining process is fully automated.

